Cytokeratin 20 (CK20)

Concentrated and Prediluted Monoclonal Antibody 901-062-041919



Catalog Number:	CM 062 A, C	PM 062 AA, H	IP 062 G10	OAI 062 T60	VLTM 062 G20
Description:	0.1, 1.0 mL, conc.	6.0, 25 mL, RTU	10 mL, RTU	60 tests, RTU	20 mL, RTU
Dilution:	1:100	Ready-to-use	Ready-to-use	Ready-to-use	Ready-to-use
Diluent:	Van Gogh Yellow	N/A	N/A	N/A	N/A

Intended Use:

For In Vitro Diagnostic Use

Cytokeratin 20 (CK20) [Ks20.8] is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of cytokeratin 20 (CK20) protein by immunohistochemistry (IHC) in formalin-fixed paraffinembedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

Cytokeratin 20 is 46 kDa intermediate filament protein that has been identified with expression primarily restricted to gastric and intestinal epithelium, urothelium, and Merkel cells. Cytokeratin 20 has been shown as a unique type I keratin that is expressed in adenocarcinomas of the colon, stomach, pancreas and bile system. It is also expressed in mucinous ovarian tumors, transitional cell carcinomas of the urinary tract, and Merkel cell carcinomas. CK20 is essentially non-reactive in squamous cell carcinomas and adenocarcinomas of the breast, lung, and endometrium, as well as nonmucinous tumors of the ovary and small cell carcinomas. Cytokeratin 20 is often used in conjunction with CK7 and other antibodies in distinguishing colon carcinomas (CK20+) from ovarian, pulmonary, and breast carcinomas.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human; others not tested

Clone: Ks20.8 Isotype: IgG2a

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: CK20

Cellular Localization: Cytoplasmic Positive Tissue Control: Colon carcinoma

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues) **Supplied As:** Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLTM062 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Peroxidase Block: Block for 5 minutes with Val Peroxidase Block.

Protocol Recommendations (VALENT Automated Slide Staining Platform) Cont'd:

Protein Block (Optional): Incubate for 10-20 minutes with Val Background

Primary Antibody: Incubate for 20 minutes.

Secondary: Incubate for 10 minutes with Val Mouse Secondary. Linker: Incubate for 10 minutes with Val Universal Linker. **Polymer:** Incubate for 10 minutes with Val Universal Polymer.

Chromogen: Incubate for 5 minutes with Val DAB.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidazed 1.

Pretreatment: Perform heat retrieval using Diva or Reveal Decloaker. Refer to the Diva or Reveal Decloaker data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe. **Polymer:** Incubate for 10-20 minutes at RT with a tertiary polymer. Chromogen: Incubate for 5 minutes at RT with Biocare's DAB - OR -

Incubate for 5-7 minutes at RT with Warp Red.

Counterstain:

Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

intelliPATH FLX Automated Slide Stainer:

IP062 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following heat retrieval.

Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

Protocol Recommendations (ONCORE™ Automated Slide Staining System):

OAI062 is intended for use with the ONCORE. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

Protocol Name: CK20

Protocol Template (Description): Ms HRP Template 1 Dewaxing (DS Option): DS2

Antigen Retrieval (AR Option): AR2, low pH; 101°C

Reagent Name, Time, Temp.: CK20, 30 min., 25°C

Protocol Recommendations (Ventana BenchMark ULTRA):

PM062 is compatible for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

Template/Detection: OptiView DAB IHC Pretreatment Protocol: CC1 64 minutes Peroxidase: Pre Primary Peroxidase Inhibitor Primary Antibody: 32 minutes, 36°C

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation

Biocare Medical 60 Berry Drive

Pacheco, CA 94553

USA



EC REP EMERGO EUROPE

Prinsessegracht 20 2514 AP The Hague

The Netherlands

Cytokeratin 20 (CK20)

Concentrated and Prediluted Monoclonal Antibody 901-062-041919



Limitations Cont'd:

times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011.

Precautions:

- 1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN3) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (3)
- 2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (4)
- 3. Microbial contamination of reagents may result in an increase in nonspecific staining.
- 4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
- 5. Do not use reagent after the expiration date printed on the vial.
- 6. The SDS is available upon request and is located at http://biocare.net.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

- 1. Perry A, Parisi JE, Kurtin PJ. Metastatic adenocarcinoma to the brain: an immunohistochemical approach. Hum Pathol. 1997 Aug;28(8):938-43.
- 2. Sack MJ, Roberts SA. Cytokeratins 20 and 7 in the differential diagnosis of metastatic carcinoma in cytologic specimens. Diagn Cytopathol. 1997 Feb;16(2):132-6.
- 3. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
- 4. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

VP Echelon Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare's antibodies by Ventana Medical Systems, Inc. Biocare and Ventana are not affiliated, associated or related in any way. Ventana®, BenchMark®, *ultra*View and OptiView are trademarks of Roche.

Rev: 062117