

Prion IHC Assay Kit B

Counterstain and Bluing Reagents Control Number: 902-IPR5033K-090617

Catalog Number: IPR 5033K G80

Description:

approximately 250 tests on the intelliPATH automated slide stainer

Intended Use:

Research Use Only. Not for use in diagnostic procedures.

Summary & Explanation:

Prion IHC Assay Kit B contains hematoxylin counterstain and bluing solution formulated for use with Prion IHC Assay Kit A (IPR5030KG15, provided separately).

Supplied As:

1. intelliPATH[™] Prion Counterstain, IPR5031G20, 4x20mL

2. intelliPATH[™] Prion Bluing Solution, IPR5032G20, 4x20mL

Storage and Stability:

Store at room temperature (22°C to 25°C) and away from light. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user.

Limitations:

This product is provided for Research Use Only (RUO) and is not for use in diagnostic procedures. Suitability for specific applications may vary and it is the responsibility of the end user to determine the appropriate application for its use.

Precautions:

1. This product contains less than 0.25% Proclin 950 and Proclin 300. Overexposure to Proclin can cause skin and eye irritation and irritation to mucous membranes and upper respiratory tract.

2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.

3. Microbial contamination of reagents may result in an increase in nonspecific staining.

4. Do not use reagent after the expiration date printed on the vial.

5. The MSDS is available upon request and is located at $\mbox{http://biocare.}$ net/support/msds/.

Technical Support:

Contact Biocare's Technical Support at 1-800-542-2002 for questions regarding this product.

Warranty:

There are no warranties, expressed or implied, which extend beyond this description. Biocare is not liable for property damage, personal injury, or economic loss caused by this product.

References:

1. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."

2. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved guideline-Third Edition CLSI document M29-A3 Wayne, PA 2005.