



RISH™ HybriSlips™
(coverslips for *in situ* hybridization procedures)

ISO
9001:2000
CERTIFIED

Control Number: 904-RI0211-050510

Catalog Number: RI0211 C

Description: 100 coverslips (24mm x 30mm) Ready To Use

Intended Use:

RISH™ HybriSlips™ are specifically designed to be used during hybridization procedures with RNA or DNA probes.

Summary & Explanation:

RISH™ HybriSlips™ are RNase free coverslips. It is recommended that users wear gloves and handle HybriSlips™ with forceps to prevent contamination by environmental nucleases.

RISH™ HybriSlips™ do not require silanization or other coating procedures to prevent nucleic acid probes from binding to slides non-specifically. Each coverslip surface is protected by a thin layer of plastic film. Either surface may be in contact with the specimen during hybridization.

Known Applications:

in situ hybridization (formalin-fixed paraffin-embedded or frozen tissue sections).

Supplied As:

100 coverslips

Storage and Stability:

Store at room temperature

Protocol Recommendations

N/A

Technical Note:

RISH™ HybriSlips™ are composed of polycarbonate (0.25mm) for rigidity and temperature resistance. RISH™ HybriSlips™ are resistant to chemicals and reagents commonly used in many laboratories.

Performance Characteristics:

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of BIOCARE products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Quality Control:

Refer NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information on Tissue Controls.

Precautions:

N/A

Troubleshooting:

Follow the reagent specific protocol recommendations according to data sheet provided. If atypical results occur, contact BIOCARE's Technical Support at 1-800-542-2002.

Limitations & Warranty:

There are no warranties, expressed or implied, which extend beyond this description. BIOCARE is not liable for property damage, personal injury, or economic loss caused by this product.

HybriSlips™ are a registered Trademark of Grace BIO-LABS

