



## Kling-On Slides

HIER Slides

Control Number: 904-SFH1103-051311

ISO  
9001:2000  
CERTIFIED**Catalog Number:** SFH1103 A, B**Description:** 1 gross, 10 gross**Intended Use:**

This product is intended for laboratory use only

**Summary & Explanation:**

For today's HIER procedures, it is necessary to use slides that have a strong positive charge to ensure tissue adherence. The Kling-On slide is indeed the slide with the strongest and most consistent charge on the market today.

Biocare quality controls each lot of Kling-Ons to ensure a minimum charge (to ensure tissue adherence) is achieved and that a maximum charge (to ensure no glass background) is not exceeded. Other factors besides charge that can affect tissue adherence include temperature of oven, type of oven (mechanical or gravity convection technology), and type of tissue.

**Kling On Benefits**

- Stronger charge than poly-L-lysine
- Most consistent charge lot to lot
- Works well in today's aggressive HIER procedures using pressure cookers.
- Frosted white portion for labeling

**Known Applications:**

Immunohistochemistry and Immunofluorescence

**Supplied As:**SFH1103A 1 gross  
SFH1103B 10 gross (1 case)**Storage and Stability:**

Store at room temperature

**Protocol Recommendations**

N/A

**Performance Characteristics:**

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

**Quality Control:**

Refer to NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information on tissue controls.

**Precautions:**

N/A

**Troubleshooting:**

Follow the reagent specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

**Limitations & Warranty:**

There are no warranties, expressed or implied, which extend beyond this description. Biocare is not liable for property damage, personal injury, or economic loss caused by this product.

