



## Thermal Test Strips

(5 Level Temperature Strips 49 - 71°C)

ISO  
9001:2000  
CERTIFIED

Control Number: 904-TS001-101810

**Catalog Number:** TS001A

**Description:** 10 tests, Ready To Use

**Intended Use:**

Temperature strips are designed to confirm that thermal testing parameters are within a desired range during assays that require specific temperature requirements.

**Summary & Explanation:**

Biocare's Thermal Test Strips can be used in a variety of applications that require specific temperature requirements (e.g. *in situ* hybridization). These adhesive temperature strips can be placed onto glass slides and run in parallel with experimental slides to ensure that the appropriate temperature range has been achieved.

**Known Applications:**

Quality control for IQ hot bars, hot plates and hybridization ovens.

**Supplied As:**

10 strips in plastic bag

**Storage and Stability:**

Store in a cool dry place below 18°C (64 °F)

**Protocol Recommendations**

Peel indicator from release paper. Apply to any clean, grease free surface (i.e., microscope slides) ensuring that the indicator has adhered to the surface before use. A color change from silver / white to black indicates that the rating has been reached.

**Performance Characteristics:**

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

**Quality Control:**

Refer to NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information on tissue controls.

**Precautions:**

N/A

**Troubleshooting:**

Follow the reagent specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

**Limitations & Warranty:**

There are no warranties, expressed or implied, which extend beyond this description. Biocare is not liable for property damage, personal injury, or economic loss caused by this product.

