

p53 Tumor Suppressor Protein (M)

Concentrated and Prediluted Monoclonal Antibody
902-042-092017

BIOCARE
M E D I C A L

Catalog Number:	ACR 042 C	APR 042 AA
Description:	1.0 ml, concentrated	6.0 ml, prediluted
Dilution:	1:100	Ready-to-use
Diluent:	Van Gogh Yellow	N/A

Intended Use:

For Research Use Only. Not for use in diagnostic procedures.

Summary and Explanation:

p53 Tumor Suppressor Protein (M) [DO-7] recognizes a 53 kDa phosphoprotein identified as the p53 suppressor gene product. It reacts with both mutant and wild type p53. Its epitope maps within the N-terminus (aa 37-45) of p53. p53 has been demonstrated in many kinds of tumors including breast, colon and prostate.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a secondary antibody is added to bind to the primary antibody. An enzyme label is then added to bind to the secondary antibody; this detection of the bound antibody is evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human; others not tested

Clone: DO-7

Isotype: IgG2b/kappa

Total Protein Concentration: ~10 mg/ml. Call for lot specific Ig concentration.

Epitope/Antigen: p53

Cellular Localization: Nuclear

Positive Tissue Control: Breast and colon carcinomas

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Staining Protocol Recommendations:

Peroxide Block: Block for 5 minutes with Biocare's Peroxidized 1.

Pretreatment: Perform heat retrieval using Biocare's Reveal or Diva Decloaker. Refer to the Reveal or Diva Decloaker product data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Biocare's Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe.

Polymer: Incubate for 10-20 minutes at RT with a tertiary polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB – OR – Incubate for 5-7 minutes at RT with Biocare's Warp Red.

Counterstain:

Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

Technical Note:

This antibody has been standardized with Biocare's MACH 4 detection system. Use TBS buffer for washing steps.

Limitations:

This product is provided for Research Use Only (RUO) and is not for use in diagnostic procedures. Suitability for specific applications may vary and it is the responsibility of the end user to determine the appropriate application for its use.

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (3)

2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come into contact with sensitive areas, wash with copious amounts of water. (4)

3. Microbial contamination of reagents may result in an increase in nonspecific staining.

4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.

5. Do not use reagent after the expiration date printed on the vial.

6. The SDS is available upon request and is located at <http://biocare.net>.

Technical Support:

Contact Biocare's Technical Support at 1-800-542-2002 for questions regarding this product.

References:

1. Suthipintawong C, *et al.* Immunostaining of estrogen receptor, progesterone receptor, MIB1 antigen, and c-erbB-2 oncoprotein in cytologic specimens: a simplified method with formalin fixation. *Diagn Cytopathol.* 1997 Aug; 17(2):127-33.

2. Alexiev BA, *et al.* Expression of c-erbB-2 oncogene and p53 tumor suppressor gene in benign and malignant breast tissue: correlation with proliferative activity and prognostic index. *Gen Diagn Pathol.* 1997 Jun; 142(5-6):271-9.

3. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."

4. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.



60 Berry Drive

Pacheco, CA 94553

USA

Rev: 062117

Tel: 800-799-9499 | www.biocare.net | Fax: 925-603-8080