

Cytokeratin LMW (8/18)

Concentrated and Prediluted Monoclonal Antibody
901-056-050619

BIOCARE
M E D I C A L

Catalog Number:	CM 056 A, C	PM 056 AA, H	IP 056 G10	OAI 056 T60	VLTM 056 G20
Description:	0.1, 1.0 mL, conc.	6.0, 25 mL, RTU	10 mL, RTU	60 tests, RTU	20 mL, RTU
Dilution:	1:100	Ready-to-use	Ready-to-use	Ready-to-use	Ready-to-use
Diluent:	Da Vinci Green	N/A	N/A	N/A	N/A

Intended Use:

For In Vitro Diagnostic Use

Cytokeratin LMW (8/18) [5D3] is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of low molecular weight cytokeratin proteins (CK 8, 18) by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

Clone 5D3 recognizes cytokeratin (CK) 8 and 18 intermediate filament proteins. These are 52.5 kDa and 45 kDa respectively. In normal tissues, 5D3 recognizes all simple and glandular epithelium. In neoplastic tissues, 5D3 may prove useful for the identification of adenocarcinomas and some squamous cell carcinomas. Use in conjunction with HMW CK to rule out squamous cell carcinoma. Keratinizing squamous carcinomas are generally negative. It works well in formalin-fixed, paraffin-embedded tissue sections.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human (does not stain mouse and rat)

Clone: 5D3

Isotype: IgG1

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: CK 8/18 (LMW)

Cellular Localization: Cytoplasmic

Positive Tissue Control: Skin

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLTM056 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

- DAB Chromogen Staining Option:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Peroxidase Block: Block for 5 minutes with Val Peroxidase Block.

Protein Block (Optional): Incubate for 10-20 minutes with Val Background Block.

Primary Antibody: Incubate for 30 minutes.

Protocol Recommendations (VALENT Automated Slide Staining Platform) Cont'd:

Secondary: Incubate for 10 minutes with Val Mouse Secondary.

Linker: Incubate for 10 minutes with Val Universal Linker.

Polymer: Incubate for 10 minutes with Val Universal Polymer.

Chromogen: Incubate for 5 minutes with Val DAB.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

- Red Chromogen Staining Option:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Protein Block (Optional): Incubate for 10-20 minutes with Val Background Block.

Primary Antibody: Incubate for 30 minutes.

Polymer: Incubate for 45 min with Val Mouse AP Polymer.

Chromogen: Incubate for 15 min with Val Fast Red.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidized 1.

Pretreatment Protocol:

Pretreatment may be performed by heat retrieval or enzyme digestion.

Heat Retrieval Method:

Perform heat retrieval using Diva or Reveal Decloaker. Refer to the Diva or Reveal Decloaker product data sheet for specific instructions.

Digestion Method:

Digest with Pepsin enzyme for 5 minutes at 37°C – OR – for 15 minutes at RT.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe.

Polymer: Incubate for 10-20 minutes at RT with a tertiary polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB – OR – Incubate for 5-7 minutes at RT with Warp Red.

Counterstain:

Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

intelliPATH FLX Automated Slide Stainer:

IP056 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following heat retrieval.

Technical Notes:

1. With cytokeratin markers, heat retrieval may provide a higher sensitivity assay; whereas, enzyme digestion may produce greater specificity. Users should validate the pretreatment method for their specific application.

2. This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

Protocol Recommendations (ONCORE™ Automated Slide Staining System):

OAI056 is intended for use with the ONCORE. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

Protocol Name: CK8/18 LMW

Cytokeratin LMW (8/18)

Concentrated and Prediluted Monoclonal Antibody
901-056-050619

BIOCARE
M E D I C A L

Protocol Template (Description): Ms HRP Template 1

Dewaxing (DS Option): DS Buffer

Antigen Retrieval (AR Option): AR2, low pH; 95°C

Reagent Name, Time, Temp.: CK8/18 LMW, 30 min., 25°C

Limitations:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN_3) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (3)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (4)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

1. Angus B, *et al.* Cytokeratins in cervical dysplasia and neoplasia: a comparative study of immunohistochemical staining using monoclonal antibodies NCL-5D3, CAM5.2, and PKK1. J Pathol. 1988 May;155(1):71-5.
2. Anbus B, *et al.* NCL-5D3: a new monoclonal antibody recognizing low molecular weight cytokeratins effective for immunohistochemistry using fixed paraffin-embedded tissue. J Pathol. 1987 Dec;153(4):377-84.
3. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
4. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.