

Placental Alkaline Phosphatase (PLAP)
Concentrated and Prediluted Rabbit Monoclonal Antibody
Control Number: 902-350-090817

Catalog Number:	ACR 350 A, C	APR 350 AA
Description:	0.1, 1.0 ml, concentrated	6.0 ml, prediluted
Dilution:	1:50-1:100	Ready-to-use
Diluent:	Van Gogh Yellow	N/A

Intended Use:

For Research Use Only. Not for use in diagnostic procedures.

Summary and Explanation:

Rabbit monoclonal (PLAP) reacts with a membrane-bound isoenzyme (Regan and Nagao type) of Placental Alkaline Phosphatase occurring in the placenta during the 3rd trimester of gestation. Studies have shown that this antibody is highly specific to PLAP and shows no crossreaction with other isoenzymes of alkaline phosphatases. It has also shown to be useful in the identification of testicular germ cell tumors. Unlike germ cell tumors, PLAP-positive somatic cell tumors uniformly express epithelial membrane antigen (EMA).

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, an enzyme labeled polymer is added to bind to the primary antibody. The detection of the bound antibody is evidenced by a colorimetric reaction.

Source: Rabbit monoclonal

Species Reactivity: Human; others not tested

Clone: SP15

Isotype: IgG

Total Protein Concentration: ~10 mg/ml. Call for lot specific Ig concentration

Epitope/Antigen: Placental Alkaline Phosphatase (PLAP)

Cellular Localization: Cell membrane

Positive Tissue Control: Placenta or seminoma

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Staining Protocol Recommendations:

Peroxide Block: Block for 5 minutes with Biocare's Peroxidized 1.

Pretreatment Solution (recommended): Diva

Pretreatment Protocol:

Heat Retrieval Method:

Retrieve sections under pressure using Biocare's Decloaking Chamber, followed by a wash in distilled water; alternatively, steam tissue sections for 45-60 minutes. Allow solution to cool for 10 minutes then wash in distilled water.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Biocare's Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: N/A

Polymer: Incubate for 30 minutes at RT with a secondary-conjugated polymer.

Chromogen:

Incubate for 5 minutes at RT with Biocare's DAB - OR - Incubate for 5-7 minutes at RT with Biocare's Warp Red.

Counterstain:

Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

Technical Note:

This antibody has been standardized with Biocare's MACH 4 detection system. Use TBS buffer for washing steps.

Limitations:

This product is provided for Research Use Only (RUO) and is not for use in diagnostic procedures. Suitability for specific applications may vary and it is the responsibility of the end user to determine the appropriate application for its use.

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (5)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (6)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Technical Support:

Contact Biocare's Technical Support at 1-800-542-2002 for questions regarding this product.

References:

1. Manivel JC, *et al.* Placental alkaline phosphatase immunoreactivity in testicular germ-cell neoplasms. *Am J Surg Pathol.* 1987-Jan; 11(1):21-9.
2. Suster S, *et al.* Germ cell tumors of the mediastinum and testis: a comparative immunohistochemical study of 120 cases. *Hum Pathol.* 1998 Jul;29(7):737-42.
3. Bailey D, *et al.* Immunohistochemical staining of germ cell tumors and intratubular malignant germ cells of the testis using antibody to placental alkaline phosphatase and a monoclonal anti-seminoma antibody. *Mod Pathol.* 1991 Mar;4(2):167-71.
4. Burke AP, Mostofi FK. Placental alkaline phosphatase immunohistochemistry of intratubular malignant germ cells and associated testicular germ cell tumors. *Hum Pathol.* 1988 Jun;19(6):663-70.
5. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
6. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.