

# Pan Cytokeratin [AE1/AE3]

Concentrated and Prediluted Antibody Cocktail  
901-011-052323

**BIOCARE**  
M E D I C A L

Available Product Formats				
Format	Catalog Number	Description	Dilution	Diluent
Concentrate	CM 011 A, B, C	0.1, 0.5, 1.0 mL	1:100	Da Vinci Green
Predilute	PM 011 AA, H	6.0, 25 mL	Ready-to-use	N/A
intelliPATH FLX	IPI 011 G10	10 mL	Ready-to-use	N/A
ONCORE Pro	OPAI 011 T60	60 tests	Ready-to-use	N/A
UltraLine – For BenchMark	VP 011 G, G25	6.0, 25 mL	Ready-to-use	N/A
Q Series– For Leica BOND-III	ALI 011 G7	7.0 mL	Ready-to-use	N/A

## Intended Use:

For In Vitro Diagnostic Use

Pan Cytokeratin [AE1/AE3] is a mouse monoclonal antibody cocktail that is intended for laboratory use in the qualitative identification of a broad spectrum of acidic and basic cytokeratin proteins by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

## Summary and Explanation:

AE1/AE3 recognizes the acidic and basic (Type I and II) subfamilies of cytokeratins. The cocktail of these two antibodies can be used to detect most human epithelia. The acidic cytokeratins have molecular weights of 56.5, 55, 51, 50, 50, 48, 46, 45, and 40 kDa. The basic cytokeratins have molecular weights of 65-67, 64, 59, 58, 56 and 52 kDa. This pan cytokeratin antibody has proved useful as a screener for the majority of human carcinomas.

## Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

**Source:** Mouse monoclonal

**Species Reactivity:** Human, mouse and rat

**Clone:** AE1/AE3

**Isotype:** IgG1

**Protein Concentration:** Call for lot specific Ig concentration.

**Epitope/Antigen:** AE1/AE3

**Cellular Localization:** Cytoplasmic

**Positive Tissue Control:** Skin or adenocarcinoma

## Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

**Supplied As:** Buffer with protein carrier and preservative

## Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

## Protocol Recommendations (intelliPATH FLX® and manual use):

**Peroxide Block:** Block for 5 minutes with Peroxidized 1.

### Pretreatment Protocol:

**Pretreatment:** Perform heat retrieval using Reveal Decloaker. Refer to the Reveal Decloaker data sheet for specific instructions.

**Protein Block (Optional):** Incubate for 5-10 minutes at RT with Background Punisher.

**Primary Antibody:** Incubate for 30 minutes at RT.

**Probe:** Incubate for 10 minutes at RT with a secondary probe.

**Polymer:** Incubate for 10-20 minutes at RT with a tertiary polymer.

**Chromogen:** Incubate for 5 minutes at RT with Biocare's DAB -OR- Incubate for 5-7 minutes at RT with Warp Red.

**Counterstain:** Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

### intelliPATH FLX Automated Slide Stainer:

IPI011 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following pretreatment.

### Technical Note:

1. With cytokeratin markers, heat retrieval may provide a higher sensitivity assay; whereas, enzyme digestion may produce greater specificity. Users should validate the pretreatment method for their specific application.

2. This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

## Protocol Recommendations (ONCORE™ Pro Automated Slide Staining System):

OPAI011 is intended for use with the ONCORE Pro. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

**Protocol Name:** Pan CK

**Protocol Template (Description):** Ms HRP Template 1

**Dewaxing (DS Buffer Option):** DS Buffer

**Antigen Retrieval (AR Option):** AR2, low pH; 90°C

**Block Option:** Buffer

**Reagent Name, Time, Temp.:** Pan CK, 30 min., 25°C

## Protocol Recommendations (Ventana BenchMark XT / ULTRA):

VP011 is intended for use with the BenchMark XT / ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

**Template/Detection:** ultraView DAB

**Pretreatment Protocol:** ULTRA CC1 Standard

**Primary Antibody:** 32 minutes, 37°C

**ultraBlock (V-Blocker BRI4001):** Incubate for 4 minutes (with appropriate Option # registered by user)

V-Blocker is recommended to be applied prior to any detection system.

# Pan Cytokeratin [AE1/AE3]

Concentrated and Prediluted Antibody Cocktail  
901-011-052323

**BIOCARE**  
M E D I C A L

## **Protocol Recommendations (Q Series – For Leica BOND-III):**

ALI011 is intended for use with the Leica BOND-III. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

**Protocol Name:** IHC Protocol F

**Detection:** Bond Polymer Refine

**HIER:** 20 min with ER1

**Peroxide Block:** 5 min

**Marker (Primary Antibody):** 15 min

**Post Primary:** 8 min

**Polymer:** 8 min

**Mixed DAB Refine:** 10 min

**Hematoxylin:** 5 min

## **Limitations:**

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

## **Quality Control:**

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2) CLSI Wayne, PA USA ([www.clsi.org](http://www.clsi.org)). 2011

## **Precautions:**

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN<sub>3</sub>) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (5)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come into contact with sensitive areas, wash with copious amounts of water. (6)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

## **Troubleshooting:**

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

## **References:**

1. Bunton TE. The immunocytochemistry of cytokeratin in fish tissues. *Vet Pathol.* 1993 Sep; 30(5):418-25.
2. Sorensen SC, *et al.* Structural distinctions among human breast epithelial cells revealed by the monoclonal antikeratin antibodies AE1 and AE3. *J Pathol.* 1987 Oct; 153(2):151-62.
3. Pinkus GS, Etheridge CL, O'Connor EM. Are keratin proteins a better tumor marker than epithelial membrane antigen? A comparative immunohistochemical study of various paraffin-embedded neoplasms using monoclonal and polyclonal antibodies. *Am J Clin Pathol.* 1986 Mar; 85(3):269-77.
4. Pinkus GS, *et al.* Optimal immunoreactivity of keratin proteins in formalin-fixed, paraffin-embedded tissue requires preliminary trypsinization. An immunoperoxidase study of various tumours using polyclonal and monoclonal antibodies. *J Histochem Cytochem.* 1985 May; 33(5):465-73.
5. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
6. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

Ultraline antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Ventana Medical Systems, Inc or Roche. Biocare, Ventana and Roche are not affiliated, associated or related in any way. Ventana®, BenchMark®, ultraView and OptiView are trademarks of Roche.

Q Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Leica Biosystems. Biocare and Leica Biosystems are not affiliated, associated or related in any way. Leica, Leica Biosystems, BOND-MAX and BOND-III are trademarks of Leica Biosystems.