

# HMB45

Concentrated and Prediluted Monoclonal Antibody  
901-057-092820

**BIOCARE**  
M E D I C A L

Available Product Formats				
Format	Catalog Number	Description	Dilution	Diluent
Concentrate	CM 057 A, B, C	0.1, 0.5, 1.0 mL	1:100	Van Gogh Yellow
Predilute	PM 057 AA	6.0 mL	Ready-to-use	N/A
intelliPATH FLX	IP 057 G10	10 mL	Ready-to-use	N/A
ONCORE	OAI 057 T60	60 tests	Ready-to-use	N/A
ONCORE Pro	OPAI 057 T60	60 tests	Ready-to-use	N/A
VALENT	VLTM 057 G20	20 mL	Ready-to-use	N/A
UltraLine – For BenchMark	AVI 057 G	6.0 mL	Ready-to-use	N/A

## Intended Use:

For In Vitro Diagnostic Use

HMB45 is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of HMB45 protein by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

## Summary and Explanation:

HMB45 reacts with a neuraminidase-sensitive oligosaccharide side chain of a glycoconjugate present in immature melanosomes. The HMB45-reactive antigen is present in cutaneous melanocytes, prenatal and infantile retinal pigment epithelium (RPE), and melanoma cells and is thought to be oncofetal in nature. This antibody has been shown to label the majority of melanomas.

## Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

**Source:** Mouse monoclonal

**Species Reactivity:** Human; others not tested

**Clone:** HMB45

**Isotype:** IgG1/kappa

**Protein Concentration:** Call for lot specific IgG concentration.

**Epitope/Antigen:** HMB45

**Cellular Localization:** Cytoplasmic

**Positive Tissue Control:** Melanoma

## Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

**Supplied As:** Buffer with protein carrier and preservative

## Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

## Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLTM057 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

## Protocol Recommendations (VALENT Automated Slide Staining Platform) Cont'd:

### - DAB Chromogen Staining Option:

**Deparaffinization:** Deparaffinize for 8 minutes with Val DePar.

**Pretreatment:** Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

**Peroxidase Block:** Block for 5 minutes with Val Peroxidase Block.

**Protein Block (Optional):** Incubate for 10-20 minutes at RT with Val Background Block.

**Primary Antibody:** Incubate for 30 minutes.

**Secondary:** Incubate for 10 minutes with Val Mouse Secondary.

**Linker:** Incubate for 10 minutes with Val Universal Linker.

**Polymer:** Incubate for 10 minutes with Val Universal Polymer.

**Chromogen:** Incubate for 5 minutes with Val DAB.

**Counterstain:** Counterstain for 5 minutes with Val Hematoxylin.

### - Red Chromogen Staining Option:

**Deparaffinization:** Deparaffinize for 8 minutes with Val DePar.

**Pretreatment:** Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

**Protein Block (Optional):** Incubate for 10-20 minutes with Val Background Block.

**Primary Antibody:** Incubate for 30 minutes.

**Polymer:** Incubate for 45 min with Val Mouse AP Polymer.

**Chromogen:** Incubate for 15 min with Val Fast Red.

**Counterstain:** Counterstain for 5 minutes with Val Hematoxylin.

## Protocol Recommendations (intelliPATH FLX® and manual use):

**Peroxide Block:** Block for 5 minutes with Peroxidized 1.

### Pretreatment Protocol:

Heat retrieval pretreatment or no pretreatment are acceptable.

Pretreatment: Perform heat retrieval using Reveal Decloaker. Refer to the Reveal Decloaker product data sheet for specific instructions.

**Protein Block (Optional):** Incubate for 5-10 minutes at RT with Background Punisher.

**Primary Antibody:** Incubate for 30 minutes at RT.

**Probe:** Incubate for 10 minutes at RT with a secondary probe.

**Polymer:** Incubate for 10-20 minutes at RT with a tertiary polymer.

**Chromogen:** Incubate for 5 minutes at RT with Biocare's DAB -OR- Incubate for 5-7 minutes at RT with Warp Red.

### Counterstain:


Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

### intelliPATH FLX Automated Slide Stainer:

IP057 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following heat retrieval.

### Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

 Biocare Medical  
60 Berry Drive  
Pacheco, CA 94553  
USA



# HMB45

Concentrated and Prediluted Monoclonal Antibody  
901-057-092820

**BIOCARE**  
M E D I C A L

## **Protocol Recommendations (ONCORE™ Automated Slide Staining System):**

OAI057 is intended for use with the ONCORE. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

**Protocol Name:** HMB45

**Protocol Template (Description):** Ms HRP Template 1

**Dewaxing (DS Option):** DS2

**Antigen Retrieval (AR Option):** AR2, low pH; 101°C

**Reagent Name, Time, Temp.:** HMB45, 30 min., 25°C

## **Protocol Recommendations (ONCORE™ Pro Automated Slide Staining System):**

OPAI057 is intended for use with the ONCORE Pro. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

**Protocol Name:** HMB45 – OR – HMB45 AP

**Protocol Template (Description):** Ms HRP Template 1 – OR – Ms AP Template 1

**Dewaxing (DS Buffer Option):** DS2-50

**Antigen Retrieval (AR Option):** AR2, low pH; 101°C

**Block Option:** Buffer

**Reagent Name, Time, Temp.:** HMB45, 30 min., 25°C

## **Protocol Recommendations (Ventana BenchMark ULTRA):**

AVI057 is intended for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

- Using **OptiView:**

**Template/Detection:** OptiView DAB IHC

**Pretreatment Protocol:** CC1 32 minutes

**Peroxidase:** Pre Primary Peroxidase Inhibitor

**Primary Antibody:** 32 minutes, 36°C

- Using **ultraView AP Red:**

**Template/Detection:** ultraView Red

**Pretreatment Protocol:** CC1 Standard

**Primary Antibody:** 32 minutes, 37°C

## **Limitations:**

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

## **Quality Control:**

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2) CLSI Wayne, PA USA (www.clsi.org). 2011

## **Precautions:**

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN<sub>3</sub>) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (2)

## **Precautions Cont'd:**

2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come into contact with sensitive areas, wash with copious amounts of water. (3)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

## **Troubleshooting:**

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

## **References:**

1. Kapur RP, *et al.* Anti-melanoma monoclonal antibody HMB45 identifies an oncofetal glycoconjugate associated with immature melanosomes. *J Histochem Cytochem.* 1992 Feb; 40(2):207-12.
2. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
3. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

VP Echelon Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare's antibodies by Ventana Medical Systems, Inc. Biocare and Ventana are not affiliated, associated or related in any way. Ventana®, BenchMark®, *ultraView* and OptiView are trademarks of Roche.