

CD138

Concentrated and Prediluted Monoclonal Antibody
901-167-052523

BIOCARE
M E D I C A L

Catalog Number:	CM 167 AK, BK, CK	PM 167 AA	IP 167 G10	AVI 167 G	VLTM 167 G20
Description:	0.1, 0.5, 1.0 mL, conc.	6.0 mL, RTU	10 mL, RTU	6.0 mL, RTU	20 mL, RTU
Dilution:	1:100	Ready-to-use	Ready-to-use	Ready-to-use	Ready-to-use
Diluent:	Van Gogh Yellow	N/A	N/A	N/A	N/A

Intended Use:

For In Vitro Diagnostic Use

CD138 [B-A38] is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of CD138 protein by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

CD138 / syndecan-1 protein backbone is a single chain molecule of 30.5 kDa. Five putative GAG attachment sites exist in the extracellular domain. GAG fine structure appears to reflect the cellular source of the syndecan. Expression of CD138 in human hematopoietic cells is restricted to plasma cells in normal bone marrow. Early B-cell precursors in human bone marrow are CD138 negative. CD138 is also expressed in endothelial cells, fibroblasts, keratinocytes and normal hepatocytes.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human; others not tested

Clone: B-A38

Isotype: IgG1

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: CD138 (Syndecan-1)

Cellular Localization: Cell membrane

Positive Tissue Control: Tonsil

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative
Van Gogh Yellow (PD902)

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLTM167 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Protocol Recommendations (VALENT® Automated Slide Staining Platform) Cont'd:

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Peroxidase Block: Block for 5 minutes with Val Peroxidase Block.

Protein Block (Optional): Incubate for 10-20 minutes with Val Background Block.

Primary Antibody: Incubate for 30 minutes.

Secondary: Incubate for 10 minutes with Val Mouse Secondary.

Linker: Incubate for 10 minutes with Val Universal Linker.

Polymer: Incubate for 10 minutes with Val Universal Polymer.

Chromogen: Incubate for 5 minutes with Val DAB.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidized 1.

Pretreatment: Perform heat retrieval using Reveal Decloaker. Refer to the Reveal Decloaker product data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe.

Polymer: Incubate for 10-20 minutes at RT with a tertiary polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB - OR - Incubate for 5-7 minutes at RT with Warp Red.

Counterstain: Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

intelliPATH™ Automated Slide Stainer:

IP167 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following heat retrieval.

Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

Protocol Recommendations (Ventana BenchMark ULTRA):

AVI167 is intended for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

Template/Detection: OptiView DAB IHC

Pretreatment Protocol: CC1 32 minutes

Peroxidase: Pre Primary Peroxidase Inhibitor

Primary Antibody: 32 minutes, 36°C

Limitations:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

CD138

Concentrated and Prediluted Monoclonal Antibody
901-167-052523

BIOCARE
M E D I C A L

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (4)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (5)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagents after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

1. Wijdenes J, *et al.* New B-cell CD antigens. BC29: CD138 (syndecan-1) workshop panel report. In: Kishimoto T, *et al.*, Eds. Leucocyte typing VI: White cell differentiation antigens. Garland Publishing Inc., New York; 1997. pp. 249-52.
2. Sun RX, *et al.* Large scale and clinical grade purification of syndecan-1+ malignant plasma cells. J Immunol Methods. 1997 June 23; 205(1):73-9.
3. Carbone A, *et al.* Reed-Sternberg cells of classical Hodgkin's disease react with the plasma cell-specific monoclonal antibody B-B4 and express human syndecan-1. Blood. 1997 May 15; 89(10):3787-94.
4. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
5. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

VP Echelon Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare's antibodies by Ventana Medical Systems, Inc. Biocare and Ventana are not affiliated, associated or related in any way. Ventana®, BenchMark®, *ultraView* and *OptiView* are trademarks of Roche.