

PAX5

Concentrated and Prediluted Monoclonal Antibody
901-207-052523

BIOCARE
M E D I C A L

Available Product Formats				
Format	Catalog Number	Description	Dilution	Diluent
Concentrate	CM 207 A, B, C	0.1, 0.5, 1.0 mL	1:100	Da Vinci Green
Predilute	PM 207 AA	6.0 mL	Ready-to-use	N/A
ONCORE	OAI 207 T60	60 tests	Ready-to-use	N/A
ONCORE Pro	OPAI 207 T60	60 tests	Ready-to-use	N/A
Q Series – For Leica BOND-III	ALI 207 G7	7.0 mL	Ready-to-use	N/A

Intended Use:

For In Vitro Diagnostic Use

PAX5 [24] is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of PAX5 protein by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

PAX5 is a B-cell specific activator protein (BSAP). In early stages of B-cell development, PAX5 influences the expression of several B-cell specific genes, such as CD19 and CD20. PAX5 is expressed primarily in pro-, pre-, and mature B-cells, but not in plasma cells. There is an excellent correlation between CD20 and PAX5 expression; however, the anti-PAX5 antibody exceeds the specificity and sensitivity of L26 (CD20) due to its expression in early B-cell differentiation and its ability to detect all committed B-cells, including classic Hodgkin's lymphoma. Studies have shown that it is very specific to B-cell lineage and does not stain T-cells. In essence, PAX5 may be a superior pan B-cell marker to CD20.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-step or two-step detection procedure can be applied. The one-step procedure will feature an enzyme-labeled polymer that binds the primary antibody. A two-step procedure will feature a linker antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human; others not tested

Clone: 24 (Previously known as BC/24)

Isotype: IgG1

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: PAX5 (B-cell)

Cellular Localization: Nuclear

Positive Tissue Control: Tonsil or B-cell lymphoma

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidized 1.

Pretreatment: Perform heat retrieval using Borg or Reveal Decloaker. Refer to the Borg or Reveal Decloaker product data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30-45 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe.

Protocol Recommendations (intelliPATH FLX and manual use)

Cont'd:

Polymer: Incubate for 10-20 minutes at RT with a tertiary polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB -OR- Incubate for 5-7 minutes at RT with Warp Red.

Counterstain:

Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

Protocol Recommendations (ONCORE™ Automated Slide Staining System):

OAI207 is intended for use with the ONCORE. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

Protocol Name: PAX-5

Protocol Template (Description): IHC Extras Template

Dewaxing (DS Option): DS Buffer

Antigen Retrieval (AR Option): AR1, high pH; 103°C

Reagent Name, Time, Temp.: PAX-5, 30 min., 25°C

- Use of **Mouse Amp HRP Detection (ORI6050)** is required for the above antibody protocol. Mouse HRP Detection (ORI6007) is not recommended.

Protocol Recommendations (ONCORE™ Pro Automated Slide Staining System):

OPAI207 is intended for use with the ONCORE Pro. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

Protocol Name: PAX-5

Protocol Template (Description): IHC Extras Template 1

Dewaxing (DS Buffer Option): DS Buffer

Antigen Retrieval (AR Option): AR1, high pH; 103°C

Block Option: Buffer

Reagent Name, Time, Temp.: PAX-5, 30 min., 25°C

- Use of **Mouse Amp HRP Detection (OPRI6050)** is required for the above antibody protocol. Mouse HRP Detection (OPRI6007) is not recommended.

Protocol Recommendations (Ventana BenchMark ULTRA):

PM207 is compatible for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

Template/Detection: OptiView DAB IHC

Pretreatment Protocol: CC1 64 minutes

Peroxidase: Pre-Primary Peroxidase Inhibitor

Primary Antibody: 32 minutes, 36°C

PAX5

Concentrated and Prediluted Monoclonal Antibody
901-207-052523

BIOCARE
M E D I C A L

Protocol Recommendations (Q Series – For Leica BOND-III):

ALI207 is intended for use with the Leica BOND-III. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

Protocol Name: IHC Protocol F
Detection: Bond Polymer Refine
HIER: 20 min with ER2
Peroxide Block: 5 min
Marker (Primary Antibody): 15 min
Post Primary: 8 min
Polymer: 8 min
Mixed DAB Refine: 10 min
Hematoxylin: 5 min

Limitations:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2) CLSI Wayne, PA, USA (www.clsi.org). 2011

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (4)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (5)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

1. Desouki MM, *et al.* Pax-5: a valuable immunohistochemical marker in the differential diagnosis of lymphoid neoplasms. *Clin Med Res.* 2010 Jul;8(2):84-8.
2. Torlakovic E, *et al.* Pax-5 expression in nonhematopoietic tissues. *Am J Clin Pathol.* 2006 Nov;126(5):798-804.
3. Torlakovic E, *et al.* The value of anti-pax-5 immunostaining in routinely fixed and paraffin-embedded sections: a novel pan pre-B and B-cell marker. *Am J Surg Pathol.* 2002 Oct;26(10):1343-50.
4. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
5. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

Ultraline antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Ventana Medical Systems, Inc or Roche. Biocare, Ventana and Roche are not affiliated, associated or related in any way. Ventana®, BenchMark®, ultraView and OptiView are trademarks of Roche.

Q Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Leica Biosystems. Biocare and Leica Biosystems are not affiliated, associated or related in any way. Leica, Leica Biosystems, BOND-MAX and BOND-III are trademarks of Leica Biosystems.