

p53

Concentrated and Prediluted Rabbit Monoclonal Antibody
901-298-111720

BIOCARE
M E D I C A L

Available Product Formats				
Format	Catalog Number	Description	Dilution	Diluent
Concentrate	CME 298 AK, BK	0.1, 0.5 mL	1:100	Van Gogh Yellow
Predilute	PME 298 AA	6.0 mL	Ready-to-use	N/A
intelliPATH FLX	IP 298 G10	10 mL	Ready-to-use	N/A
ONCORE Pro	OPAI 298 T60	60 tests	Ready-to-use	N/A
VALENT	VLTR 298 G20	20 mL	Ready-to-use	N/A
UltraLine – For BenchMark	AVI 298 G	6.0 mL	Ready-to-use	N/A

Intended Use:

For In Vitro Diagnostic Use

p53 [EP9] is a rabbit monoclonal antibody that is intended for laboratory use in the qualitative identification of p53 protein by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

p53 acts both as a tumor-suppressor and transcription factor. Activation by DNA damage or other stress signals triggers cell-cycle arrest, apoptosis and DNA repair. The nuclear p53 gene is located on chromosome 17p, a frequent site of allele loss in many tumors (60%) including breast, colon and lung. Studies have shown this high affinity p53 rabbit monoclonal is very specific and is superior to other p53 mouse monoclonal antibodies. This antibody recognizes both wild-type and mutant p53.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-step or two-step detection procedure can be applied. A one-step procedure will feature an enzyme labeled polymer that binds the primary antibody. A two-step procedure will feature a linker antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Rabbit monoclonal

Species Reactivity: Human; others not tested

Clone: EP9 (previously known as Y5)

Isotype: IgG

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: p53

Cellular Localization: Nuclear

Positive Tissue Control: Breast and colon carcinomas

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative
Van Gogh Yellow (PD902)

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLTR298 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Peroxidase Block: Block for 5 minutes with Val Peroxidase Block.

Protein Block: Incubate for 10 minutes with Val Background Block.

Primary Antibody: Incubate for 30 minutes.

Secondary: N/A

Linker: Incubate for 10 minutes with Val Universal Linker.

Polymer: Incubate for 20 minutes with Val Universal Polymer.

Chromogen: Incubate for 5 minutes with Val DAB.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidized 1.

Pretreatment: Perform heat retrieval using Diva or Reveal Decloaker. Refer to the Diva or Reveal Decloaker product data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30 minutes at RT.

Probe: N/A

Polymer: Incubate for 20-30 minutes at RT with a secondary-conjugated polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB -OR- Incubate for 5-7 minutes at RT with Warp Red.

Counterstain: Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

intelliPATH FLX Automated Slide Stainer:

IP298 is intended for use with the intelliPATH FLX. Refer to the User Manual for specific instructions for use. When using the intelliPATH FLX, peroxide block with intelliPATH FLX Peroxidase Blocking Reagent (IPB5000) may be performed following heat retrieval.

Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

p53

Concentrated and Prediluted Rabbit Monoclonal Antibody
901-298-111720

BIOCARE
M E D I C A L

Protocol Recommendations (ONCORE™ Pro Automated Slide Staining System):

OPA1298 is intended for use with the ONCORE Pro. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Editor should be programmed as follows:

Protocol Name: p53 Rb

Protocol Template (Description): Rb HRP Template 1

Dewaxing (DS Buffer Option): DS2-50

Antigen Retrieval (AR Option): AR2, low pH; 101°C

Block Option: Buffer

Reagent Name, Time, Temp.: p53 Rb, 30 min., 25°C

Protocol Recommendations (Ventana BenchMark ULTRA):

AVI298 is intended for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

Template/Detection: OptiView DAB IHC

Pretreatment Protocol: CC1 32 minutes

Peroxidase: Pre Primary Peroxidase Inhibitor

Primary Antibody: 32 minutes, 36°C

Limitations:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2) CLSI Wayne, PA USA (www.clsi.org). 2011

Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (3)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (4)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

1. Harris CC. p53: at the crossroads of molecular carcinogenesis and risk assessment. *Science*. 1993 Dec 24;262(5142):1980-1.
2. Alexiev BA, *et al.* Expression of c-erbB-2 oncogene and p53 tumor suppressor gene in benign and malignant breast tissue: correlation with proliferative activity and prognostic index. *Gen Diagn Pathol*. 1997 Jun;142(5-6):271-9.
3. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
4. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

Produced using Abcam's RabMab® technology. RabMab® technology is covered by the following U.S. Patents, No. 5,675,063 and/or 7,429,487.

VP Echelon Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare's antibodies by Ventana Medical Systems, Inc. Biocare and Ventana are not affiliated, associated or related in any way. Ventana®, BenchMark®, *ultraView* and OptiView are trademarks of Roche.