

# CD47 [EPR21794]

Concentrated and Prediluted Rabbit Monoclonal Antibody  
902-3284-062722

**BIOCARE**  
M E D I C A L

Available Product Formats				
Format	Catalog Number	Description	Dilution	Diluent
Concentrate	ACR 3284 A, C	0.1, 1.0 mL	1:100	Da Vinci Green
Predilute	APR 3284 AA	6.0 mL	Ready-to-use	N/A
UltraLine – For BenchMark	AVR 3284 G	6.0 mL	Ready-to-use	N/A
Q Series– For Leica BOND-III	ALR 3284 G7	7.0 mL	Ready-to-use	N/A

## Intended Use:

For Research Use Only. Not for use in diagnostic procedures.

## Summary and Explanation:

CD47 has been shown to have a role in both cell adhesion by acting as an adhesion receptor for THBS1 on platelets, and in the modulation of integrins.<sup>1</sup> In the brain, it appears to play an important role in memory formation and synaptic plasticity in the hippocampus.<sup>2</sup> CD47 is a receptor for SIRPA, binding to which prevents maturation of immature dendritic cells and inhibits cytokine production by mature dendritic cells.<sup>3</sup> Interaction with SIRPG mediates cell-cell adhesion, enhances superantigen-dependent T-cell-mediated proliferation and costimulates T-cell activation.<sup>1</sup> CD47 may play a role in membrane transport and/or integrin dependent signal transduction. CD47 has been described as broadly expressed in many human tissues, with abundant expression in epithelial cells, the brain, and in ovarian cancer.<sup>4</sup>

## Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-step or two-step detection procedure can be applied. A one-step procedure will feature an enzyme labeled polymer that binds the primary antibody. A two-step procedure will feature a linker antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

**Source:** Rabbit monoclonal

**Species Reactivity:** Human; other species not tested

**Clone:** [EPR21794]

**Isotype:** IgG

**Protein Concentration:** Call for lot specific Ig concentration.

**Epitope/Antigen:** CD49

**Cellular Localization:** Cell membrane

**Positive Tissue Control:** Liver cancer, placenta

## Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

**Supplied As:** Buffer with protein carrier and preservative

## Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

## Staining Protocol Recommendations (intelliPATH FLX® and manual use):

**Peroxide Block:** Block for 5 minutes with Peroxidized 1.

**Pretreatment:** Perform heat retrieval using Diva Decloaker. Refer to the Diva Decloaker data sheet for specific instructions.

**Protein Block (Optional):** Incubate for 5-10 minutes at RT with Background Punisher.

**Primary Antibody:** Incubate for 30 minutes at RT.

**Probe:** N/A

**Polymer:** Incubate for 30 minutes at RT with a tertiary polymer.

**Chromogen:** Incubate for 5 minutes at RT with Biocare's DAB – OR – Incubate for 5-7 minutes at RT with Warp Red.

**Counterstain:** Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

## Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

## Staining Protocol Recommendations (Ventana BenchMark ULTRA):

AVR3284 is intended for use with the BenchMark ULTRA. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

**Template/Detection:** OptiView DAB IHC

**Pretreatment Protocol:** CC1 64 minutes

**Peroxidase:** Pre-Primary Peroxidase Inhibitor

**Primary Antibody:** 48 minutes, 36°C

## Staining Protocol Recommendations (Q Series – For Leica BOND-III):

ALR3284 is intended for use with the Leica BOND-III. Refer to the User Manual for specific instructions for use. Recommended protocol parameters are as follows:

**Protocol Name:** IHC Protocol F

**Detection:** Bond Polymer Refine

**HIER:** 30 min with ER2

**Peroxide Block:** 5 min

**Marker (Primary Antibody):** 15 min

**Post Primary:** 8 min

**Polymer:** 8 min

**Mixed DAB Refine:** 10 min

**Hematoxylin:** 5 min

## Limitations:

This product is provided for Research Use Only (RUO) and is not for use in diagnostic procedures. Suitability for specific applications may vary and it is the responsibility of the end user to determine the appropriate application for its use.

## Precautions:

1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN<sub>3</sub>) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976)<sup>5</sup>

2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come into contact with sensitive areas, wash with copious amounts of water.<sup>6</sup>

3. Microbial contamination of reagents may result in an increase in nonspecific staining.

4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.

5. Do not use reagent after the expiration date printed on the vial.

6. The SDS is available upon request and is located at <http://biocare.net>.

## Technical Support:

Contact Biocare's Technical Support at 1-800-542-2002 for questions regarding this product.

 Biocare Medical

60 Berry Drive

Pacheco, CA 94553

USA

TP v1 (10/26/2021)

Tel: 800-799-9499 | [www.biocare.net](http://www.biocare.net) | Fax: 925-603-8080

## CD47 [EPR21794]

Concentrated and Prediluted Rabbit Monoclonal Antibody  
902-3284-062722

**BIOCARE**  
M E D I C A L

### References:

1. Piccio L, Vermi W, Boles KS, et al. Adhesion of human T cells to antigen-presenting cells through SIRP $\beta$ 2-CD47 interaction costimulates T-cell proliferation. *Blood* 2005; 105 (6); 2421-2427.
2. Miyashita M, Ohnishi H, Okazawa H, et al. Promotion of Neurite and Filopodium Formation by CD47: Roles of Integrins, Rac, and Cdc42. *Mol Biol Cell* 2004; 15(8); 3950-3963.
3. Latour S, Tanaka H, Demeure C, et al. Bidirectional negative regulation of human T and dendritic cells by CD47 and its cognate receptor signal-regulator protein-alpha: down-regulation of IL-12 responsiveness and inhibition of dendritic cell activation. *J Immunol* 2001; 167(5); 2547-2554.
4. Soto-Pantoja DR, Kaur S, and Roberts DD. CD47 signaling pathways controlling cellular differentiation and responses to stress. *Crit Rev Biochem Mol Biol* 2015; 50(3); 212-230.
5. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
6. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.

Ultraline antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Ventana Medical Systems, Inc or Roche. Biocare, Ventana and Roche are not affiliated, associated or related in any way. Ventana®, BenchMark®, ultraView and OptiView are trademarks of Roche.

Q Series antibodies are developed solely by Biocare Medical LLC and do not imply approval or endorsement of Biocare antibodies by Leica Biosystems. Biocare and Leica Biosystems are not affiliated, associated or related in any way. Leica, Leica Biosystems, BOND-MAX and BOND-III are trademarks of Leica Biosystems.



Biocare Medical

60 Berry Drive

Pacheco, CA 94553

USA

TP v1 (10/26/2021)

Tel: 800-799-9499 | [www.biocare.net](http://www.biocare.net) | Fax: 925-603-8080