

HPV Type 6 Probe

Label(s): Digoxigenin

Volume: 500µl

SKU: HPV6-500-D-A

Analyte Specific Reagent. Analytical and performance characteristics are not established.

Storage and Handling: Store probe at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date.

Material Provided:

- HPV Type 6 Probe (40ng/µl).

Intended Use:

Empire Genomics' HPV Type 6 Probe is a digoxigenin labeled cDNA probe designed to hybridize HPV Type 6 E6/E7 sequences.

Principle of Method: Chromogenic *in situ* hybridization (CISH)

Warning and Precautions: Product does not contain any human or animal components. See safety data sheet for more detailed safety and handling information. Contains formamide in low concentration. Do not use expired probe. Do not reuse probe. Avoid cross-contamination. Read instructions for use in full before use.

- H360d May damage the unborn child.

- P309+P310 IF exposed or if you feel unwell:

Immediately call a POISON CENTER or doctor/physician.

- P201 Obtain special instructions before use.

- H315+H320 Causes skin and eye irritation.

- P262 Do not get in eyes, on skin, or on clothing.

- P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

- P501 Dispose of contents/container to local/regional/national/international regulations.

- P280 Wear protective gloves/protective clothing/eye protection/face protection.

